
# Galltec Mess- und Regeltechnik GmbH D-71145 Bondorf · Germany Tel. +49 (0)7457-9453-0 · Fax +49 (0)7457-3758 E-Mail: sensoren@galltec.de · Internet:www.galltec-mela.de

#### MELA Sensortechnik GmbH

D-07987 Mohlsdorf-Teichwolframsdorf · Germany Tel. +49(0)3661-62704-0 · Fax +49(0)3661-62704-20 E-mail:mela@melasensor.de · Internet: www.galltec-mela.de





with one changeover contact, scale range 30...100%rh, IP00

#### Application


**Technical Data** 

The hygro module HM80 is a humidity-dependent switch that can be fitted in equipment such as hygrostats, humidifiers, dehumidifiers, ventilating fans, driers and many more. The module represents an on-off controller with changeover contact. The switch connection is via a connecting terminal, but can also be supplied ready-made with cable connections. Several versions of different lengths are available as a shaft. Protection of the module is of the IP00 type.

#### **Description of the Hygrostat**

The humidity measuring element, produced by Galltec under the name Polyga®, consists of several synthetic fabric bands each with 90 individual fibres with a diameter of 3µm. A special process gives the fibre hygroscopic properties. The measuring element adsorbs and desorbs humidity. The swelling effect, which is predominantly in a lengthways direction, is carried via a suitable lever system to a microswitch with an extremely small switching path. The measuring element reacts quickly and precisely to the change in air humidity. By adjusting the setpoint value control knob, the lever system is engaged so that when the set air humidity is reached the microswitch is activated.

The fan shaped measuring element is protected by a perforated sensor tube. The hygro module is designed for pressureless systems. The mounting position should be chosen such that condensed water cannot get onto the microswitch.



point value Fw, contact 1/2 opens and contact 1/4 closes.

CE

| scale range   |  |  |  |  |
|---|--|--|--|--|
| range of operation  |  |  |  |  |
| measuring medium air, pressureless, non-aggressive            |  |  |  |  |
| switching difference (microswitch) ref. to 50%rh approx. 4%rh |  |  |  |  |
| switching distance between the microswitches                  |  |  |  |  |
| for HM80-2 0+15%rh  |  |  |  |  |
| breaking capacity of the changeover contact                   |  |  |  |  |
| ohmic load (cos φ=1)15A AC 230V                               |  |  |  |  |
| inductive load (cos φ=0,7) 2A AC 230V                         |  |  |  |  |
| direct voltage 0.25A DC 230V                                  |  |  |  |  |
| low voltage100mA, 125V AC                                     |  |  |  |  |
| contact material silver                                       |  |  |  |  |
| allowable storage temperature3060°C                           |  |  |  |  |
| allowable operating temperature                               |  |  |  |  |
| medium temp. coefficient0.2%/K relative to 20°C and 50%rh     |  |  |  |  |
| adjustment at average air pressure 430 m NN                   |  |  |  |  |
| allowable air speed   |  |  |  |  |
| with gauze protection (order no. 20.014) 15m/sec              |  |  |  |  |
| time constant T <sub>63</sub> at v=2m/sec                     |  |  |  |  |

sensor length, material ...... 220mm, high-grade steel

connecting terminals ...... for conductor cross sections 1.5mm² cable connection ...... via twist nipple M20x1.5

applied directives / standards

low-voltage directive 2014/35/EU EMC directive 2014/30/EU

DIN EN 60730-1:2012-10 DIN EN 60730-2-13:2008-09

type of protection (external adjusting knob) ...... IP00 measuring element .....

...... Polyga®-measuring element, water resistent, washable weight ....... approx. 0.25 kg

This information is based on current knowledge and is intended to provide details of our products and their possible applications. It does not, therefore, act as a guarantee of specific properties of the products described or of their suitability for a particular application. It is our experience that the equipment may be used across a broad spectrum of applications under the most varied conditions and loads. We cannot appraise every individual case. Purchasers and/or users are responsible for checking the equipment for suitability for any particular application. Any existing industrial rights of protection must be observed. The perfect quality of our products is guaranteed under our General Conditions of Sale. Issue: December 2015 HM80\_E. Subject to modifications.

### Operating instructions for channel hygrostat HM80 and HM80-2


#### Mounting

- The hygrostats must not come into direct contact with water (e.g. splashed water when cleaning the climatic chamber etc.)
- The mounting location should be chosen so that a representative measurement of the air humidity can be guaranteed, i.e. the humidity readings at the mounting location should correspond to those in the room as far as possible.
- The hygrostat should be exposed to the flow of air.

#### Operating information:

Note that, with restrictions in the upper range of operation, the possible tolerances (measurement accuracy, switching difference and temperature coefficient) should be observed when adjusting the switch point.

#### Preferred mounting positions



#### **Maintenance**

The measuring element is maintenance-free in pure ambient air. Depending on their type and concentration, aggressive media containing solvents can cause incorrect readings or cause the humidistat to fail. Substances deposited on the measuring element (e.g. resin aerosols, paint aerosols, smoke deposits etc.) are harmful as they eventually form a water-repellent film. The water-resistant property of the Galltec measuring elements allows cleaning to be carried out in water. Solvents cannot be used for this purpose. A light-duty detergent is recommended, but any residue should always be washed out thoroughly.

A special process ensures that Galltec sensors have good long-term stability. Regeneration is not necessary, but is also not harmful.

Contact with the inner parts of the humidistat nullifies the warranty.

#### Calibration

Equipment with Galltec hygrostats is correctly set by the factory at a room temperature of 23°C and 50% rel. humidity, relative to the average air pressure of 430m NN.

If, however, subsequent adjustment should be necessary, the following procedure should be observed.

- Ensure that the ambient humidity and the ambient temperature are constant.
- If possible, use a psychrometer for checking (no chekking equipment with capacitive sensors).
- Leave the equipment to be checked for at least 1 hour in a constant checking climate.
- The adjuster screw is at the end of the sensor fixed with screw securing lacquer. After removing the lacquer, the adjuster screw can be moved. A right-hand rotation means that the measured value goes down, and with a left-hand rotation the measured value goes up. After calibration, the adjuster screw should again be secured.

#### Note:

Moving the adjuster screw nullifies the guarantee.

**Important.** The water absorption capacity of the air is influenced, amongst other things, by the temperature. This is a physical law (which can be seen from the hx diagram of Mollier). The higher the air temperature, the greater the volume of water vapour that can absorbed up to saturation point (100%rh). If a hygrostat is now calibrated at fluctuating air temperature, there is an irregular, non-homogeneous measured medium and there are automatically calibration errors. The table below shows the influence of the air temperature on air humidity. If, for example, calibration occurs at an air temperature of 20 °C and 50%rh, and at a temperature fluctuation of just ±1 °K, then there will be a humidity fluctuation in the measured medium (air) of ±3.2%rh.

|       | 10°C      | 20°C      | 30°C      | 50°C      |
|-------|-----------|-----------|-----------|-----------|
| 10%rh | +/-0,7%rh | +/-0,6%rh | +/-0,6%rh | +/-0,5%rh |
| 50%rh | +/-3,5%rh | +/-3,2%rh | +/-3,0%rh | +/-2,6%rh |
| 90%rh | +/-6,3%rh | +/-5,7%rh | +/-5,4%rh | +/-4,6%rh |

#### **Maintenance instructions**

for humidity measuring equipment with Polyga®humidity measuring element.

#### Impact of dirt

The measuring element is maintenance-free in pure ambient air. Depending on their type and concentration, aggressive media containing solvents can cause incorrect readings or cause the humidistat to fail. Hygroscopic humidity measuring elements - and these include particularly the capacitive measuring elements, resite measuring elements and fibre measuring elements (Polyga) - are sensitive if a water-repellent film forms on the surface of the elements. Such sensors and hygrostats cannot be used for example during wood drying as, depending on the type of wood to be dried, resin aerosols in the surrounding air are deposited on the measuring element. The same applies to lacquer drying equipment where there are paint aerosols in the surrounding air.

The water-repellent property of the Galltec® humidity measuring element allows cleaning to be carried out in water. An important benefit when the sensors are used in extreme atmospheres.

#### Cleaning instruction

for humidity sensors, hygrostats in channel design as well as all shaft equipment, FG80.., TFG80.., HG80, HG80-2, HM80 and HM80-2.

The humidity measuring equipment is designed with a perforated sensor tube. The humidity and temperature measuring elements are in the interior of the sensor tube. The humidity measuring element is arranged axially and the temperature measuring element is seated sideways at the top or bottom between the perforation holes.

Measuring equipment with Pt100 glass measuring resistors can be immersed in water. Other temperature sensors, in particular semi-conductor sensors or customer-specific temperature sensors should not come into contact with water. Enquire if in doubt.

The measuring elements are designed for use in pressureless air (gases). The measurement accuracy depends on the degree of pollution of the element. The humidity measuring element, in particular, loses its hygroscopic properties if the surface is covererd with grease, soot, smoke deposits, paint, resinous substances etc. By cleaning the elements, their function can be reproduced, but only if no damage is caused by acids, alkaline solutions or other aggressive substances.

### Cleaning process

#### 1. Disconnect the device from the power supply!

2. Dip the sensor tube into a receptacle containing clean water (20°C) and, with a gentle rotating motion, disperse the dirt deposits. If the dirt contains grease deposits, it is recommended that a mild detergent be added to the water.

Do not brush or treat with any other cleaning utensils. Only the sensor tube should be immersed - not the housing.

The sensor tube is open to the housing interior (0.8mm hole)

- 3. As mild detergents are known to contain chemical substances, rinse carefully after cleaning. Cleaning residue will impair the measured result.
- 4. Air drying. Where a measuring element is moistened with water, the device indicates 100% relative humidity. If necessary, it is possible to carry out sensitive recalibration at the adjusting spindle at the end of the sensor. This should only be done where there are large deviations. Slight movement of the adjusting spindle of a wet element causes the measured value in the dry area to be badly out. Here, there is an intensifying effect of the linearisation (factor 6). Indications of 98..100% relative humidity at the wet element are adequate.

The accuracy in the dry area must be determined under normal climatic conditions.

## The measuring element must not be dried using warm or hot air (hair dryer).

Cleaning should be carried out for no longer than some seconds.

#### Installation advises

On installation ensure that there is sufficient overcurrent protection (e. g. fuse). Also a separation device (e. g. plug or switch) has to be installed.

Further detailes informationen you will find in "HUMIDITY SENSOR IN ACCORDANCE WITH THE ABSORPTION PRINCIPLE"